

■Concise communications

Survey on Sterilization Validation Practice in Japanese Hospitals

Chie Takeuchi*, Hiroyoshi Kobayashi*

Introduction

Adequate sterilization of medical instruments is one of the important practices to prevent healthcare associated infections. Japanese private study group “The Forum on Infection Prevention and Medical Instruments” founded in April 1995, conducted nationwide survey on sterilization and sterilization assurance practice among 500 hospitals in 1998¹⁾ and 2002²⁾ after the publication of the first guideline from Japanese Society of Medical Instrumentation (JSMI) for sterilization assurance in healthcare settings in 2000³⁾. The guideline revised in 2005⁴⁾, and after then the third nationwide survey was conducted in the same manner in 2007 and was reported⁵⁾. As revised guideline in one of the questionnaires the implementation of the sterilization process validation in hospitals was asked to reply. As a result, 207 hospitals (45.4%) responded that sterilization process validation was implemented. To investigate the details of the practices of sterilization process validation and its responsibility in these 207 hospitals, an additional survey was conducted.

1. Method

In the third survey, a questionnaire was sent to Japanese hospitals with over 400 beds or which certified sterilization technicians were belonging to. 456 hospitals replied and 207 (45.4%) had implemented sterilization process validation. Additional questionnaire was prepared to investigate the details of the practices of sterilization process validation recommended in JSMI

guideline published in 2005⁵⁾ and its responsibility. The questionnaire was sent to those 207 hospitals.

2. Result

79 hospitals replied to the additional questionnaires sent to 207 hospitals. The reply ratio is 38.2%. Survey results are shown in Table 1 ~ 3. Installation qualifications (IQ) are fairly well implemented in most hospitals. In each item of IQ, number of hospitals where it is not performed is approximately less than 30 % except some items. 78.9% of steam sterilizers, and 74.3% of ethylene oxide gas (EOG) sterilizers and also 74.3% of hydrogen peroxide gas plasma sterilizers (plasma) had been installed before the publication of the guideline 2005.

Operational qualifications (OQ) are well implemented except the check of the temperature and pressure variation in the chamber, identification of the cold spot, biological indicator (BI) test at the points other than the cold spot. Leak test is not performed in almost 35% of the hospitals for all three kinds of sterilizers.

Implementations of performance qualifications (PQ) showed poorly low compliance rates compared to IQ and OQ.

Although in many hospitals all qualifications were performed by hospital sterilization technicians, rather large number of hospitals has requested to be performed by manufacturers and suppliers.

* Tokyo Healthcare University of Postgraduate School

Table 1. Validation items of Installation Qualification (IQ)

| Validation items | Sterilization method | No of facilities replied | Performed by | | | | | | | | | |
|--|----------------------|--------------------------|--------------------|-------|-------------|-------|--------------------------|-------|--------|-------|---------------|-------|
| | | | Hospital personnel | | Outsourcing | | Manufacturer or supplier | | Others | | Not performed | |
| 1. Specification for the accessories of the sterilizer and air conditioning | Steam | 73 | 34 | 46.6% | 2 | 2.7% | 37 | 50.7% | 1 | 1.4% | 3 | 4.1% |
| | EOG | 68 | 24 | 35.3% | 6 | 8.8% | 34 | 50.0% | 4 | 5.9% | 3 | 4.4% |
| | Plasma | 46 | 11 | 23.9% | 1 | 2.2% | 28 | 60.9% | 2 | 4.3% | 7 | 15.2% |
| 2. Sterilization process programme | Steam | 75 | 30 | 40.0% | 8 | 10.7% | 38 | 50.7% | 1 | 1.3% | 1 | 1.3% |
| | EOG | 69 | 24 | 34.8% | 10 | 14.5% | 34 | 49.3% | 3 | 4.3% | 1 | 1.4% |
| | Plasma | | | | | | | | | | | |
| 3. Aeration time | Steam | | | | | | | | | | | |
| | EOG | 68 | 35 | 51.5% | 11 | 16.2% | 23 | 33.8% | 3 | 4.4% | 0 | 0.0% |
| | Plasma | | | | | | | | | | | |
| 4. Inquiry of MSDS | Steam | | | | | | | | | | | |
| | EOG | 61 | 32 | 52.5% | 8 | 13.1% | 12 | 19.7% | 6 | 9.8% | 3 | 4.9% |
| | Plasma | 52 | 24 | 46.2% | 7 | 13.5% | 11 | 21.2% | 5 | 9.6% | 5 | 9.6% |
| 5. Storage condition for sterilant | Steam | | | | | | | | | | | |
| | EOG | 65 | 35 | 53.8% | 10 | 15.4% | 9 | 13.8% | 6 | 9.2% | 5 | 7.7% |
| | Plasma | 59 | 25 | 42.4% | 9 | 15.3% | 12 | 20.3% | 3 | 5.1% | 10 | 16.9% |
| 6. Manual for booster use | Steam | | | | | | | | | | | |
| | EOG | | | | | | | | | | | |
| | Plasma | 58 | 15 | 25.9% | 7 | 12.1% | 19 | 32.8% | 3 | 5.2% | 15 | 25.9% |
| 7. Material compatibility verification of medical device/instrument for plasma sterilization | Steam | | | | | | | | | | | |
| | EOG | | | | | | | | | | | |
| | Plasma | 59 | 26 | 44.1% | 12 | 20.3% | 15 | 25.4% | 5 | 8.5% | 5 | 8.5% |
| 8. Material compatibility verification of packaging material for plasma sterilization | Steam | | | | | | | | | | | |
| | EOG | | | | | | | | | | | |
| | Plasma | 59 | 24 | 40.7% | 12 | 20.3% | 17 | 28.8% | 6 | 10.2% | 5 | 8.5% |
| 9. Calibration of measurement devices | Steam | 73 | 18 | 24.7% | 0 | 0.0% | 39 | 53.4% | 2 | 2.7% | 16 | 21.9% |
| | EOG | 68 | 13 | 19.1% | 4 | 5.9% | 32 | 47.1% | 5 | 7.4% | 15 | 22.1% |
| | Plasma | 60 | 12 | 20.0% | 1 | 1.7% | 34 | 56.7% | 3 | 5.0% | 12 | 20.0% |
| 10. Quality of the steam | Steam | 74 | 37 | 50.0% | 2 | 2.7% | 14 | 18.9% | 4 | 5.4% | 19 | 25.7% |
| | EOG | | | | | | | | | | | |
| | Plasma | | | | | | | | | | | |
| 11. Specification for the physical monitors used for routine process control | Steam | 74 | 21 | 28.4% | 4 | 5.4% | 45 | 60.8% | 1 | 1.4% | 7 | 9.5% |
| | EOG | 65 | 12 | 18.5% | 8 | 12.3% | 37 | 56.9% | 3 | 4.6% | 7 | 10.8% |
| | Plasma | 57 | 10 | 17.5% | 3 | 5.3% | 35 | 61.4% | 2 | 3.5% | 10 | 17.5% |
| 12. Manual for Bowie & Dick test | Steam | 73 | 52 | 71.2% | 14 | 19.2% | 7 | 9.6% | 0 | 0.0% | 5 | 6.8% |
| | EOG | | | | | | | | | | | |
| | Plasma | | | | | | | | | | | |
| 13. Leak test procedure | Steam | 73 | 17 | 23.3% | 3 | 4.1% | 30 | 41.1% | 3 | 4.1% | 20 | 27.4% |
| | EOG | 65 | 8 | 12.3% | 6 | 9.2% | 31 | 47.7% | 5 | 7.7% | 15 | 23.1% |
| | Plasma | 57 | 5 | 8.8% | 2 | 3.5% | 31 | 54.4% | 4 | 7.0% | 15 | 26.3% |
| 14. Maintenance of sterilizer | Steam | 74 | 34 | 45.9% | 1 | 1.4% | 37 | 50.0% | 2 | 2.7% | 3 | 4.1% |
| | EOG | 66 | 22 | 33.3% | 6 | 9.1% | 33 | 50.0% | 4 | 6.1% | 3 | 4.5% |
| | Plasma | 60 | 22 | 36.7% | 2 | 3.3% | 30 | 50.0% | 2 | 3.3% | 7 | 11.7% |
| 15. Installation of sterilizer | Steam | 73 | 35 | 47.9% | 3 | 4.1% | 34 | 46.6% | 4 | 5.5% | 3 | 4.1% |
| | EOG | 66 | 27 | 40.9% | 7 | 10.6% | 28 | 42.4% | 6 | 9.1% | 1 | 1.5% |
| | Plasma | 61 | 24 | 39.3% | 2 | 3.3% | 27 | 44.3% | 4 | 6.6% | 8 | 13.1% |
| 16. Documentation for IQ | Steam | 72 | 28 | 38.9% | 4 | 5.6% | 27 | 37.5% | 4 | 5.6% | 12 | 16.7% |
| | EOG | 64 | 21 | 32.8% | 8 | 12.5% | 20 | 31.3% | 5 | 7.8% | 12 | 18.8% |
| | Plasma | 59 | 16 | 27.1% | 5 | 8.5% | 20 | 33.9% | 6 | 10.2% | 15 | 25.4% |
| 17. Record for IQ | Steam | 72 | 29 | 40.3% | 10 | 13.9% | 22 | 30.6% | 3 | 4.2% | 10 | 13.9% |
| | EOG | 64 | 25 | 39.1% | 11 | 17.2% | 15 | 23.4% | 4 | 6.3% | 10 | 15.6% |
| | Plasma | 59 | 18 | 30.5% | 9 | 15.3% | 14 | 23.7% | 5 | 8.5% | 14 | 23.7% |
| 18. Installation after the publication of the guideline in 2005 | Steam | 38 | 4 | 10.5% | 1 | 2.6% | 2 | 5.3% | 1 | 2.6% | 30 | 78.9% |
| | EOG | 35 | 4 | 11.4% | 2 | 5.7% | 2 | 5.7% | 1 | 2.9% | 26 | 74.3% |
| | Plasma | 35 | 5 | 14.3% | 1 | 2.9% | 1 | 2.9% | 2 | 5.7% | 26 | 74.3% |
| 19. Others | Steam | 6 | 2 | 33.3% | 0 | 0.0% | 1 | 16.7% | 1 | 16.7% | 2 | 33.3% |
| | EOG | 5 | 0 | 0.0% | 1 | 20.0% | 1 | 20.0% | 1 | 20.0% | 2 | 40.0% |
| | Plasma | 4 | 0 | 0.0% | 0 | 0.0% | 0 | 0.0% | 2 | 50.0% | 2 | 50.0% |

Table 2. Validation items of Operational Qualification (OQ)

| Validation items | Sterilization method | No of facilities replied | Performed by | | | | | | | | | |
|---|----------------------|--------------------------|--------------------|-------|-------------|-------|--------------------------|-------|--------|-------|---------------|--------|
| | | | Hospital personnel | | Outsourcing | | Manufacturer or supplier | | Others | | Not performed | |
| 1. Bowie & Dick test | Steam | 78 | 40 | 51.3% | 31 | 39.7% | 6 | 7.7% | 0 | 0.0% | 6 | 7.7% |
| | EOG | | | | | | | | | | | |
| | Plasma | | | | | | | | | | | |
| 2. Leak test | Steam | 75 | 15 | 20.0% | 5 | 6.7% | 26 | 34.7% | 1 | 1.3% | 28 | 37.3% |
| | EOG | 65 | 9 | 13.8% | 8 | 12.3% | 21 | 32.3% | 4 | 6.2% | 23 | 35.4% |
| | Plasma | 61 | 7 | 11.5% | 4 | 6.6% | 26 | 42.6% | 3 | 4.9% | 21 | 34.4% |
| 3. Check of the minimum pressure during air removal phase | Steam | 77 | 38 | 49.4% | 19 | 24.7% | 17 | 22.1% | 1 | 1.3% | 7 | 9.1% |
| | EOG | 67 | 28 | 41.8% | 18 | 26.9% | 12 | 17.9% | 3 | 4.5% | 8 | 11.9% |
| | Plasma | 60 | 17 | 28.3% | 14 | 23.3% | 19 | 31.7% | 3 | 5.0% | 12 | 20.0% |
| 4. Check of the number of pulsations during air removal phase | Steam | 78 | 37 | 47.4% | 21 | 26.9% | 17 | 21.8% | 1 | 1.3% | 7 | 9.0% |
| | EOG | | | | | | | | | | | |
| | Plasma | | | | | | | | | | | |
| 5. Check of steam supply | Steam | 78 | 35 | 44.9% | 22 | 28.2% | 18 | 23.1% | 1 | 1.3% | 7 | 9.0% |
| | EOG | | | | | | | | | | | |
| | Plasma | | | | | | | | | | | |
| 6. Check of the temperature during sterilization phase | Steam | 78 | 43 | 55.1% | 26 | 33.3% | 10 | 12.8% | 0 | 0.0% | 4 | 5.1% |
| | EOG | 69 | 32 | 46.4% | 23 | 33.3% | 7 | 10.1% | 3 | 4.3% | 6 | 8.7% |
| | Plasma | 61 | 20 | 32.8% | 16 | 26.2% | 12 | 19.7% | 3 | 4.9% | 11 | 18.0% |
| 7. Check of the pressure during sterilization phase | Steam | 78 | 41 | 52.6% | 24 | 30.8% | 9 | 11.5% | 1 | 1.3% | 7 | 9.0% |
| | EOG | 69 | 32 | 46.4% | 21 | 30.4% | 7 | 10.1% | 3 | 4.3% | 7 | 10.1% |
| | Plasma | 61 | 21 | 34.4% | 18 | 29.5% | 12 | 19.7% | 3 | 4.9% | 10 | 16.4% |
| 8. Check of the time(duration) of sterilization phase | Steam | 76 | 37 | 48.7% | 25 | 32.9% | 11 | 14.5% | 0 | 0.0% | 7 | 9.2% |
| | EOG | 70 | 32 | 45.7% | 24 | 34.3% | 7 | 10.0% | 3 | 4.3% | 7 | 10.0% |
| | Plasma | 61 | 18 | 29.5% | 19 | 31.1% | 13 | 21.3% | 3 | 4.9% | 11 | 18.0% |
| 9. Check of the aeration time | Steam | | | | | | | | | | | |
| | EOG | 55 | 30 | 54.5% | 22 | 40.0% | 2 | 3.6% | 1 | 1.8% | 2 | 3.6% |
| | Plasma | | | | | | | | | | | |
| 10. Door safety (not open at the pressure other than the atmospheric) | Steam | 70 | 35 | 50.0% | 24 | 34.3% | 12 | 17.1% | 1 | 1.4% | 3 | 4.3% |
| | EOG | | | | | | | | | | | |
| | Plasma | | | | | | | | | | | |
| 11. Check of the temperature variation in the chamber | Steam | 64 | 5 | 7.8% | 2 | 3.1% | 6 | 9.4% | 2 | 3.1% | 50 | 78.1% |
| | EOG | 58 | 5 | 8.6% | 2 | 3.4% | 1 | 1.7% | 2 | 3.4% | 48 | 82.8% |
| | Plasma | 50 | 2 | 4.0% | 0 | 0.0% | 0 | 0.0% | 4 | 8.0% | 44 | 88.0% |
| 12. Check of the pressure variation in the chamber | Steam | 59 | 4 | 6.8% | 1 | 1.7% | 3 | 5.1% | 2 | 3.4% | 50 | 84.7% |
| | EOG | 52 | 3 | 5.8% | 2 | 3.8% | 0 | 0.0% | 2 | 3.8% | 45 | 86.5% |
| | Plasma | 48 | 2 | 4.2% | 0 | 0.0% | 0 | 0.0% | 4 | 8.3% | 42 | 87.5% |
| 13. Identification of the cold point | Steam | 71 | 18 | 25.4% | 10 | 14.1% | 14 | 19.7% | 6 | 8.5% | 24 | 33.8% |
| | EOG | 67 | 18 | 26.9% | 8 | 11.9% | 11 | 16.4% | 6 | 9.0% | 25 | 37.3% |
| | Plasma | 58 | 9 | 15.5% | 5 | 8.6% | 12 | 20.7% | 7 | 12.1% | 25 | 43.1% |
| 14. BI test at the cold point | Steam | 77 | 35 | 45.5% | 21 | 27.3% | 8 | 10.4% | 2 | 2.6% | 18 | 23.4% |
| | EOG | 71 | 30 | 42.3% | 20 | 28.2% | 5 | 7.0% | 2 | 2.8% | 19 | 26.8% |
| | Plasma | 63 | 18 | 28.6% | 18 | 28.6% | 8 | 12.7% | 4 | 6.3% | 19 | 30.2% |
| 15. CI test at the cold point | Steam | 77 | 35 | 45.5% | 21 | 27.3% | 7 | 9.1% | 2 | 2.6% | 18 | 23.4% |
| | EOG | 71 | 30 | 42.3% | 20 | 28.2% | 4 | 5.6% | 2 | 2.8% | 19 | 26.8% |
| | Plasma | 63 | 19 | 30.2% | 17 | 27.0% | 7 | 11.1% | 4 | 6.3% | 19 | 30.2% |
| 16. BI test at the points other than the cold point | Steam | 75 | 20 | 26.7% | 12 | 16.0% | 7 | 9.3% | 2 | 2.7% | 36 | 48.0% |
| | EOG | 68 | 19 | 27.9% | 13 | 19.1% | 4 | 5.9% | 3 | 4.4% | 30 | 44.1% |
| | Plasma | 61 | 12 | 19.7% | 10 | 16.4% | 6 | 9.8% | 4 | 6.6% | 30 | 49.2% |
| 17. CI test at the point other than the cold point | Steam | 76 | 30 | 39.5% | 16 | 21.1% | 8 | 10.5% | 2 | 2.6% | 23 | 30.3% |
| | EOG | 69 | 27 | 39.1% | 16 | 23.2% | 4 | 5.8% | 3 | 4.3% | 21 | 30.4% |
| | Plasma | 62 | 16 | 25.8% | 14 | 22.6% | 6 | 9.7% | 4 | 6.5% | 24 | 38.7% |
| 18. Documentation for OQ | Steam | 76 | 40 | 52.6% | 8 | 10.5% | 8 | 10.5% | 1 | 1.3% | 22 | 28.9% |
| | EOG | 69 | 32 | 46.4% | 12 | 17.4% | 4 | 5.8% | 3 | 4.3% | 21 | 30.4% |
| | Plasma | 62 | 27 | 43.5% | 5 | 8.1% | 7 | 11.3% | 3 | 4.8% | 22 | 35.5% |
| 19. Record of OQ | Steam | 75 | 36 | 48.0% | 23 | 30.7% | 8 | 10.7% | 0 | 0.0% | 13 | 17.3% |
| | EOG | 69 | 27 | 39.1% | 24 | 34.8% | 4 | 5.8% | 2 | 2.9% | 14 | 20.3% |
| | Plasma | 60 | 19 | 31.7% | 18 | 30.0% | 8 | 13.3% | 2 | 3.3% | 15 | 25.0% |
| 20. Others | Steam | 7 | 2 | 28.6% | 1 | 14.3% | 1 | 14.3% | 0 | 0.0% | 3 | 42.9% |
| | EOG | 6 | 1 | 16.7% | 1 | 16.7% | 1 | 16.7% | 0 | 0.0% | 3 | 50.0% |
| | Plasma | 4 | 0 | 0.0% | 0 | 0.0% | 0 | 0.0% | 0 | 0.0% | 4 | 100.0% |

Table 3. Validation items of Performance Qualification (PQ)

| Validation items | Sterilization method | No of facilities replied | Performed by | | | | | | | | | |
|---|----------------------|--------------------------|--------------------|-------|-------------|-------|--------------------------|-------|--------|-------|---------------|--------|
| | | | Hospital personnel | | Outsourcing | | Manufacturer or supplier | | Others | | Not performed | |
| 1. Definition of reference load | Steam | 66 | 20 | 30.3% | 9 | 13.6% | 4 | 6.1% | 1 | 1.5% | 34 | 51.5% |
| | EOG | 62 | 15 | 24.2% | 10 | 16.1% | 4 | 6.5% | 3 | 4.8% | 32 | 51.6% |
| | Plasma | 54 | 8 | 14.8% | 7 | 13.0% | 5 | 9.3% | 4 | 7.4% | 31 | 57.4% |
| 2. Loading contents of reference load | Steam | 67 | 24 | 35.8% | 7 | 10.4% | 5 | 7.5% | 3 | 4.5% | 29 | 43.3% |
| | EOG | 63 | 17 | 27.0% | 8 | 12.7% | 4 | 6.3% | 3 | 4.8% | 32 | 50.8% |
| | Plasma | 55 | 12 | 21.8% | 8 | 14.5% | 4 | 7.3% | 5 | 9.1% | 27 | 49.1% |
| 3. Loading configuration of reference load | Steam | 67 | 24 | 35.8% | 8 | 11.9% | 6 | 9.0% | 3 | 4.5% | 27 | 40.3% |
| | EOG | 63 | 19 | 30.2% | 9 | 14.3% | 4 | 6.3% | 3 | 4.8% | 29 | 46.0% |
| | Plasma | 55 | 14 | 25.5% | 8 | 14.5% | 4 | 7.3% | 5 | 9.1% | 25 | 45.5% |
| 4. Temperature measurement inside the reference load | Steam | 65 | 11 | 16.9% | 1 | 1.5% | 5 | 7.7% | 2 | 3.1% | 46 | 70.8% |
| | EOG | 61 | 9 | 14.8% | 3 | 4.9% | 4 | 6.6% | 2 | 3.3% | 43 | 70.5% |
| | Plasma | 54 | 8 | 14.8% | 0 | 0.0% | 4 | 7.4% | 4 | 7.4% | 38 | 70.4% |
| 5. Definition of Process Challenge Device(PCD) | Steam | 65 | 20 | 30.8% | 13 | 20.0% | 3 | 4.6% | 3 | 4.6% | 29 | 44.6% |
| | EOG | 61 | 18 | 29.5% | 12 | 19.7% | 2 | 3.3% | 4 | 6.6% | 28 | 45.9% |
| | Plasma | 53 | 12 | 22.6% | 8 | 15.1% | 3 | 5.7% | 5 | 9.4% | 27 | 50.9% |
| 6. Temperature measurement inside the PCD | Steam | 64 | 3 | 4.7% | 2 | 3.1% | 6 | 9.4% | 4 | 6.3% | 49 | 76.6% |
| | EOG | 59 | 3 | 5.1% | 5 | 8.5% | 3 | 5.1% | 3 | 5.1% | 45 | 76.3% |
| | Plasma | 52 | 3 | 5.8% | 1 | 1.9% | 4 | 7.7% | 5 | 9.6% | 39 | 75.0% |
| 7. Temperature measurement at multiple points in the chamber | Steam | 59 | 4 | 6.7% | 1 | 1.7% | 2 | 3.3% | 2 | 3.3% | 51 | 85.0% |
| | EOG | 56 | 3 | 5.4% | 3 | 5.4% | 1 | 1.8% | 1 | 1.8% | 48 | 85.7% |
| | Plasma | 48 | 2 | 4.2% | 0 | 0.0% | 1 | 2.1% | 4 | 8.3% | 41 | 85.4% |
| 8. Pressure measurement at multiple points in the chamber | Steam | 56 | 2 | 3.6% | 0 | 0.0% | 1 | 1.8% | 2 | 3.6% | 51 | 91.1% |
| | EOG | 52 | 3 | 5.8% | 2 | 3.8% | 0 | 0.0% | 1 | 1.9% | 46 | 88.5% |
| | Plasma | 45 | 2 | 4.4% | 0 | 0.0% | 0 | 0.0% | 4 | 8.9% | 39 | 86.7% |
| 9. Identification of cold point in the reference load | Steam | 63 | 13 | 20.6% | 6 | 9.5% | 8 | 12.7% | 5 | 7.9% | 31 | 49.2% |
| | EOG | 60 | 12 | 20.0% | 6 | 10.0% | 7 | 11.7% | 4 | 6.7% | 31 | 51.7% |
| | Plasma | 51 | 8 | 15.7% | 4 | 7.8% | 5 | 9.8% | 7 | 13.7% | 27 | 52.9% |
| 10. BI test at the point identified by 9. | Steam | 65 | 21 | 32.3% | 14 | 21.5% | 3 | 4.6% | 1 | 1.5% | 29 | 44.6% |
| | EOG | 61 | 17 | 27.9% | 13 | 21.3% | 2 | 3.3% | 1 | 1.6% | 30 | 49.2% |
| | Plasma | 53 | 13 | 24.5% | 10 | 18.9% | 4 | 7.5% | 4 | 7.5% | 23 | 43.4% |
| 11. CI test at the point identified by 9. | Steam | 65 | 24 | 36.9% | 14 | 21.5% | 3 | 4.6% | 1 | 1.5% | 26 | 40.0% |
| | EOG | 61 | 21 | 34.4% | 12 | 19.7% | 2 | 3.3% | 1 | 1.6% | 27 | 44.3% |
| | Plasma | 53 | 15 | 28.3% | 9 | 17.0% | 4 | 7.5% | 4 | 7.5% | 22 | 41.5% |
| 12. BI test at cold point in PCD | Steam | 65 | 23 | 35.4% | 13 | 20.0% | 4 | 6.2% | 1 | 1.5% | 27 | 41.5% |
| | EOG | 61 | 21 | 34.4% | 14 | 23.0% | 2 | 3.3% | 0 | 0.0% | 27 | 44.3% |
| | Plasma | 53 | 13 | 24.5% | 7 | 13.2% | 4 | 7.5% | 4 | 7.5% | 25 | 47.2% |
| 13. CI test at cold point in PCD | Steam | 64 | 25 | 39.1% | 13 | 20.3% | 4 | 6.3% | 1 | 1.6% | 24 | 37.5% |
| | EOG | 61 | 23 | 37.7% | 13 | 21.3% | 2 | 3.3% | 0 | 0.0% | 26 | 42.6% |
| | Plasma | 53 | 13 | 24.5% | 7 | 13.2% | 4 | 7.5% | 4 | 7.5% | 25 | 47.2% |
| 14. BI test at the point other than cold point | Steam | 65 | 18 | 27.7% | 10 | 15.4% | 3 | 4.6% | 2 | 3.1% | 35 | 53.8% |
| | EOG | 61 | 15 | 24.6% | 10 | 16.4% | 1 | 1.6% | 2 | 3.3% | 35 | 57.4% |
| | Plasma | 52 | 10 | 19.2% | 7 | 13.5% | 4 | 7.7% | 4 | 7.7% | 28 | 53.8% |
| 15. CI test at the point other than cold point | Steam | 64 | 22 | 34.4% | 12 | 18.8% | 3 | 4.7% | 2 | 3.1% | 29 | 45.3% |
| | EOG | 60 | 20 | 33.3% | 11 | 18.3% | 1 | 1.7% | 2 | 3.3% | 29 | 48.3% |
| | Plasma | 51 | 13 | 25.5% | 9 | 17.6% | 4 | 7.8% | 4 | 7.8% | 23 | 45.1% |
| 16. BI test at the point other than cold point in the PCD | Steam | 63 | 13 | 20.6% | 9 | 14.3% | 3 | 4.8% | 3 | 4.8% | 37 | 58.7% |
| | EOG | 59 | 13 | 22.0% | 10 | 16.9% | 2 | 3.4% | 1 | 1.7% | 35 | 59.3% |
| | Plasma | 51 | 8 | 15.7% | 8 | 15.7% | 5 | 9.8% | 4 | 7.8% | 27 | 52.9% |
| 17. CI test at the point other than cold point in the PCD | Steam | 64 | 14 | 21.9% | 9 | 14.1% | 3 | 4.7% | 3 | 4.7% | 37 | 57.8% |
| | EOG | 60 | 13 | 21.7% | 10 | 16.7% | 2 | 3.3% | 1 | 1.7% | 36 | 60.0% |
| | Plasma | 52 | 8 | 15.4% | 8 | 15.4% | 5 | 9.6% | 4 | 7.7% | 28 | 53.8% |
| 18. Setup of the sterilization condition by overkill method | Steam | 50 | 6 | 12.0% | 4 | 8.0% | 8 | 16.0% | 0 | 0.0% | 33 | 66.0% |
| | EOG | 47 | 3 | 6.4% | 3 | 6.4% | 7 | 14.9% | 0 | 0.0% | 34 | 72.3% |
| | Plasma | 44 | 4 | 9.1% | 3 | 6.8% | 7 | 15.9% | 3 | 6.8% | 28 | 63.6% |
| 19. Setup of the sterilization condition by half-cycle method | Steam | 49 | 1 | 2.0% | 1 | 2.0% | 4 | 8.2% | 0 | 0.0% | 43 | 87.8% |
| | EOG | 47 | 3 | 6.4% | 3 | 6.4% | 5 | 10.6% | 0 | 0.0% | 36 | 76.6% |
| | Plasma | 43 | 1 | 2.3% | 1 | 2.3% | 4 | 9.3% | 2 | 4.7% | 35 | 81.4% |
| 20. Dryness of the reference load after sterilization | Steam | 61 | 29 | 47.5% | 17 | 27.9% | 2 | 3.3% | 1 | 1.6% | 14 | 23.0% |
| | EOG | 65 | 28 | 43.1% | 17 | 26.2% | 2 | 3.1% | 1 | 1.5% | 19 | 29.2% |
| | Plasma | 57 | 23 | 40.4% | 14 | 24.6% | 1 | 1.8% | 0 | 0.0% | 20 | 35.1% |
| 21. Packaging integrity of the reference load after sterilization | Steam | 65 | 28 | 43.1% | 17 | 26.2% | 2 | 3.1% | 1 | 1.5% | 19 | 29.2% |
| | EOG | 57 | 23 | 40.4% | 14 | 24.6% | 1 | 1.8% | 0 | 0.0% | 20 | 35.1% |
| | Plasma | 50 | 17 | 34.0% | 12 | 24.0% | 1 | 2.0% | 3 | 6.0% | 17 | 34.0% |
| 22. Documentation of PQ | Steam | 66 | 21 | 31.8% | 8 | 12.1% | 6 | 9.1% | 3 | 4.5% | 29 | 43.9% |
| | EOG | 60 | 17 | 28.3% | 9 | 15.0% | 3 | 5.0% | 3 | 5.0% | 29 | 48.3% |
| | Plasma | 53 | 14 | 26.4% | 5 | 9.4% | 4 | 7.5% | 5 | 9.4% | 25 | 47.2% |
| 23. Record of PQ | Steam | 66 | 22 | 33.3% | 13 | 19.7% | 7 | 10.6% | 2 | 3.0% | 25 | 37.9% |
| | EOG | 60 | 19 | 31.7% | 13 | 21.7% | 3 | 5.0% | 2 | 3.3% | 24 | 40.0% |
| | Plasma | 53 | 14 | 26.4% | 10 | 18.9% | 4 | 7.5% | 4 | 7.5% | 22 | 41.5% |
| 24. Others | Steam | 2 | 1 | 50.0% | 0 | 0.0% | 0 | 0.0% | 0 | 0.0% | 1 | 50.0% |
| | EOG | 2 | 1 | 50.0% | 0 | 0.0% | 0 | 0.0% | 0 | 0.0% | 1 | 50.0% |
| | Plasma | 2 | 0 | 0.0% | 0 | 0.0% | 0 | 0.0% | 0 | 0.0% | 2 | 100.0% |

3. Discussion

After the first survey in 1998, the first Japanese guideline for sterilization assurance in healthcare settings was published in 2000³⁾ from JSMI and this guideline was revised in 2005 to broaden its subject from only sterilization assurance program to cleaning, packaging, sterilization, sterilization assurance⁴⁾. And this revised JSMI guideline recommends sterilization process validation. As one of the other activities of JSMI, the certification program of the second class sterilization technician (Certified Sterilization Service Technician: CSST) and the first class sterilization technician (Certified Sterilization Specialist: CSS) were established by JSMI in 2000 and 2003, respectively. As of August 2008, 2,543 sterilization technicians are certified as the second class (CSST), and 124 sterilization technicians are certified as the first class (CSS).

This time the survey has been performed to investigate the impact of Japanese guideline and certification system on the actual practice in healthcare

settings. In the results, most of IQ and OQ items in many hospitals responded are fairly well performed. However implementation of PQ in many hospitals is still insufficient. In order to obtain the better quality of sterilization in Japanese hospitals, the strategies to improve the compliance of PQ should be the main subjects.

This study was supported by The Forum on Infection Prevention and Medical Instruments.

■Reference

- 1) Kobayashi H, Nagai I. Survey on sterilization assurance activities in Japanese healthcare facilities. *Jpn J Hosp Supply* 1999; 4 (1) : 40-52. (in Japanese)
- 2) Kobayashi H, Nagai I. Survey on sterilization assurance activities in Japanese healthcare facilities II. *JJMI* 2003; 73:114-123. (in Japanese)
- 3) JSMI. *Guideline for Sterility Assurance in Healthcare Setting*, 2000. Tokyo: JSMI 2000. (in Japanese)
- 4) JSMI. *Guideline for Sterility Assurance in Healthcare Setting*, 2005. Ed: Kobayashi H. Tokyo: JSMI 2005. (in Japanese)
- 5) Takeuchi C, Kobayashi H. Survey on Sterilization Assurance Activities in Japanese Healthcare Facilities III. *JJMI* 2008; 78:433-445. (in Japanese)