Research Brief

## A study of the Ultrasonic Cleaning Process Indicator for the Washer-Disinfector with the Ultrasonic Cleaning System

Toshiaki Shimizu 1,2

<sup>1</sup> Tokyo Healthcare University Postgraduate School <sup>2</sup> SAKURA SEIKI Co., Ltd

**Background:** At the Washer-Disinfector with the Ultrasonic Cleaning system (WDUC) used to reprocessing surgical instruments, the aluminium foil test is used in order to detect the ultrasonic energy level in cleaning solution of  $WDUC^{1,2)}$ . The erosion on the aluminium foil shows ultrasonic energy level. However, the aluminium foil test does not serve as a test with a process indicator of WDUC. Furthermore, preparation of the aluminium foil test are requires time, and the powder of aluminum foil which is generated by the result of erosion contaminates  $WDUC^{3}$ . Accordingly it is desirable to provide alternative test device.

**Objective:** Creating the practical Ultrasonic Cleaning Process Indicator (UCPI) for WDUC by using the aluminium foil. Creating trial piece of UCPI for WDUC meets following requirements: 1) Possible to check ultrasonic energy level after all process including high temperature drying cycle of WDUC. 2) Not contaminate WDUC by the powder of aluminum foil. 3) Can be installed easily in the tray of WDUC.

**Method:** Aluminium foil was heat-sealed into polypropylene bag with propylene glycol as the trial piece of UCPI shown in Figure 1. The polypropylene band was attached and hooked at the tray of WDUC to keep vertical. Then the performance test of an experimental product of the process indicator was carried out by using a real WDUC. Precondition of WDUC is shown on Table 1.

The bubble to keep an indicator vertical in the cleaning solution.

Bag : Polypropylene, width is 25mm, length is 40mm, thickness is 0.06mm. Contents: Propylene glycol. (KANTO CHEMICAL CO., INC.)



Figure 1. A trial piece.

Aluminium foil: 20mm in width, 30mm in length, 0.012mm in thickness (My foil, UACJ Foil Corporation Co., Ltd.)

The band to install indicator to the tray of WDUC: material is polypropylene.

Apparatus	SAKURA Automatic Ultrasonic and Jet Washer WUS II -3100DXW, SAKURA SEIKI Co., Ltd.		
Program	Cycle name	Temperature	Time
	Preliminary washing	10 degree C	2min.
	Jet washing	45 degree C	5min.
	Ultrasonic washing	45 degree C	10min.
	Note: Frequency is 40kHz. Power is 2400W. Volume is 220L.		
	Rinse	40 degree C	1min.
	Disinfection	90 degree C	5min.
	Drying	110 degree C	10min.

Table 1. Precondition of WDUC.

**Result:** After the performance test, pattern of the erosion on the aluminium foil was observed as shown in Figure 2 and also all of above issues were solved. The UCPI did not damage after ultrasonic washing cycle (45 degree C, 10 minutes) and high temperature drying cycle (110 degree C, 10 minutes) of WDUC. Additionally, it was observed that powder of aluminum foil was staying in of UCPI as shown Figure 2.



Figure 2. After the performance test.

**Discussions:** The trial piece of UCPI for WDUC was possible to check ultrasonic energy level after all process including high temperature drying cycle of WDUC. Also it did not contaminate WDUC by the powder of aluminum foil, and could be installed easily in the tray. Additionally this trial piece of UCPI can also be used for the ultrasonic cleaner.

**Conclusion:** This trial piece of the UCPI was confirmed to successfully detect the ultrasonic energy efficacy after processing with WDUC. Additionally, it is one-touch attachable to a tray of WDUC. Also it can be stuck in a scrapbook to be stored as a record.

## ■Reference

- Department of Health (UK). Decontamination of surgical instruments (CFPP 01-01). https://www.gov.uk/government/uploads/system/uploads/attachment\_data/file/148522/CFPP\_01-01D\_Final.pdf. accessed December 10, 2014.
- Pfeifer M. Validation of SonoCheck for the Monitoring of Ultrasonic Energy of Ultrasonic Cleaner. http://www.healthmark.info/CleaningVerification/SonoCheck/ValidationofSonoCheck.pdf. accessed November 16, 2014.
- 3) Zwahlen A, Wild MD, Jung C. Comparison of Methods for Testing Ultrasound in the Cleaning Bath: DAGA 2014 Oldenburg 716-17.